

Digital Chlorine Dioxide (ClO₂) Sensor



1. Overview

The AGRINOVO-CLO2-100 is a digital chlorine dioxide (ClO₂) electrode for continuous monitoring of disinfectant residual in water. It integrates the measuring electronics and microprocessor inside the electrode body, with built-in temperature measurement and automatic temperature and pH compensation. The sensor communicates over RS485 Modbus-RTU and supports two-point (zero and slope) calibration. The isolated power and signal design and PPS housing make it suitable for demanding water treatment environments.

Key Features

- Chlorine dioxide range: 0-20 mg/L
- Resolution: 0.01 mg/L
- Accuracy: ± 0.1 mg/L
- Built-in temperature measurement
- Two-point calibration (zero + slope)
- Automatic temperature and pH compensation
- RS485 Modbus-RTU interface
- 9-27 VDC, ≤ 0.5 W

Applications

- Drinking water disinfection monitoring
 - Water treatment and distribution
 - Cooling water and cooling tower treatment
 - Aquaculture and RAS water management
 - Food and beverage process water
 - Wastewater disinfection control
-

2. Specifications

Parameter	Specification
Measurement Principle	Electrochemical electrode
Chlorine Dioxide Range	0.00-20.00 mg/L
Resolution	0.01 mg/L
Accuracy	±0.1 mg/L
Temperature Range	-10.0 to 110.0 °C
Temperature Resolution	0.1 °C
Temperature Accuracy	±0.5 °C
Temperature Element	TH10K thermistor
Temperature Compensation	Automatic / Manual
Communication	RS485, Modbus-RTU
Isolation	2500 Vrms
Supply Voltage	9-27 VDC
Power Consumption	≤0.5W
Housing Material	PPS
Mounting Thread	M20×1.5 or G3/4

3. Wiring

Wire Color	Function	Description
Red	V+	Power Supply (9-27V DC)
Black	GND	Power Ground
Green	RS485-A	Data+ (T/R+)
White	RS485-B	Data- (T/R-)

4. Communication Settings

Parameter	Value
Protocol	Modbus-RTU
Baud Rate	9600 bps (configurable)
Data Bits	8
Parity	None
Stop Bits	1
Default Address	0x01

Supported baud rates: 1200, 2400, 4800, 9600, 19200 bps.

5. Register Map

Measurement values are available in two formats. Function 0x03 returns IEEE-754 float32 (each value spans 2 registers, transmitted low word first). Function 0x04 returns scaled 16-bit integers, each followed by a register carrying the decimal places and unit code.

Measurement Registers (Function 0x03 / 0x04 read)

Address	Description	Range	Data Type
0x0000-0x0001	Chlorine Dioxide (ClO ₂)	0.00-20.00 mg/L	Float32
0x0004-0x0005	Electrode Signal	-10.00 to 40.00 mV	Float32
0x0008-0x0009	Temperature	-10.0 to 110.0 °C	Float32

Configuration Registers (Function 0x03 read / 0x06 write)

Address	Description	Default	Range / Notes
0x0019	Calibration Status	Not calibrated	Bit0 = zero point, Bit1 = slope point
0x001A	Electrode Offset	0.00 mg/L	±0.30 mg/L (value ×100)
0x001C	Electrode Slope	100.0%	30.0-300.0% (value ×10)
0x001E	Device Address	1	1-247 (255 = broadcast)
0x001F	Baud Rate Code	3 (9600)	0=1200, 1=2400, 2=4800, 3=9600, 4=19200
0x0020	Temp Compensation Type	1 (auto)	0 = manual, 1 = automatic
0x0021	Temp Setpoint / Offset	25.0 °C / 0.0 °C	value ×10 (see note)
0x0023	pH Compensation Switch	1 (on)	0 = off, 1 = on
0x0024	pH Compensation Value	4.00 pH	0.00-14.00 pH (value ×100)
0x002D	Filter Coefficient	—	1-16

Register 0x0021 holds the manual temperature setpoint when 0x0020 = manual, or the temperature offset when 0x0020 = automatic. Example: a stored value of 250 means 25.0 °C; to set 10.0 °C write 100 (0x0064).

Information Registers (Function 0x03 read, 0x06 write where applicable)

Address	Description	Notes
0x0040	Work Mode	0x10 = measure, 0x50 = setting, 0x60 = calibration
0x0041	Mode Parameter	Mode-specific
0x0043	Calibration Control / Status	Write standard value to calibrate (see Section 8)
0x0046	Software Version	Read only
0x0047	Hardware Version	Read only
0x0048-0x0049	Serial Number	Read only

6. Reading Data

Read Chlorine Dioxide value (Float32 from 0x0000-0x0001)

Read 2 registers starting at 0x0000:

Request: 01 03 00 00 00 02 [CRC]
Response: 01 03 04 E7 2F 41 1F [CRC]

Decoding (word-swap example):

Received bytes E7 2F 41 1F → reorder the two 16-bit words to 41 1F E7 2F →
IEEE-754 float = 9.99 mg/L

Read all measurement values

Read 10 registers from 0x0000:

Request: 01 03 00 00 00 0A C5 CD

Returns Chlorine Dioxide, Electrode Signal, and Temperature as float32 values.

Integer format (Function 0x04)

Function 0x04 returns each value as a scaled integer followed by its decimal/unit register. For example, a value of 998 with 2 decimal places represents **9.98 mg/L**; a temperature value of 250 with 1 decimal place represents **25.0 °C**.

7. Address and Baud Rate Configuration

Change Address (0x01 → 0x02)

Write the new address to register 0x001E:

```
01 06 00 1E 00 02 [CRC]
```

Change Baud Rate (to 2400)

Write the baud rate code to register 0x001F:

```
01 06 00 1F 00 01 [CRC]
```

Broadcast Discovery

Use address 0xFF (255) with only one device connected.

8. Calibration

The sensor supports two-point calibration: a slope point and a zero point. Standard solutions are used to calibrate a newly installed sensor or to re-calibrate periodically. Calibration is performed by writing to the calibration control register 0x0043 with function 0x06.

Procedure

1. Immerse the electrode in the standard solution.
2. Wait for the reading to stabilize.

3. Write the calibration command to register 0x0043.
4. Read register 0x0043 to check the result:
 - 0 = calibration successful (returned to measure mode)
 - 1 = calibrating (re-read status shortly)
 - 2 = no valid standard value received within 180s
 - 3 = signal could not stabilize or value out of range within 180s
 - 4 = electrode performance (slope or offset) out of allowed range
5. Repeat for the second point.

Slope Point

Write the standard solution value as an integer scaled $\times 100$. For a 5.00 mg/L standard, write 500 (0x01F4):

```
01 06 00 43 01 F4 [CRC]
```

Zero Point

Write the zero point code 0x0001:

```
01 06 00 43 00 01 [CRC]
```

Clear Calibration

Write 0x7FFF to register 0x0043 to clear all calibration data:

```
01 06 00 43 7F FF [CRC]
```

9. Factory Reset

A factory reset clears calibration data and restores temperature compensation settings to default. It requires two commands.

Enter Setting Mode

Write 0x0050 to the work mode register 0x0040:

01 06 00 40 00 50 88 22

Send Reset Command

Write 0x7FFF to the mode parameter register 0x0041:

01 06 00 41 7F FF B9 AE

The sensor restores defaults and restarts.

10. Installation Notes

Safety

- Do not hot-plug
- Verify wiring before applying power
- Mount using the M20×1.5 or G3/4 thread

Sensor Care

- Keep the measurement tip submerged
- Avoid air bubbles at the electrode tip
- Clean with distilled water; no bare hand contact
- Re-calibrate periodically with fresh standards

AGRINOVO Modular IoT sensors and controllers for agriculture and aquaculture.

Request a quote: agrinovo.io/contact · Full range: agrinovo.io/products

contact@agrinovo.io · WhatsApp **+972 54 688 8148**