

Digital Conductivity Sensor



1. Overview

The AGRINOVO-EC-120 is a digital conductivity sensor for measuring electrical conductivity (EC) and temperature in water. It supports multiple electrode constants for different measurement ranges and includes slope/intercept calibration.

Key Features

- EC measurement ($\mu\text{S}/\text{cm}$)
- Temperature measurement
- Accuracy: $\pm 1.0\%$ FS
- Multiple K values
- Slope/intercept calibration
- Auto temp compensation
- RS485 Modbus-RTU
- Low power: $\leq 0.5\text{W}$

Applications

- Water quality monitoring
- Hydroponics
- Aquaculture
- Agricultural irrigation

2. Specifications

Parameter	Specification
Accuracy	$\pm 1.0\%$ FS
Temp Compensation	0–60°C (automatic)
Power Consumption	$\leq 0.5\text{W}$
Supply Voltage	12-24 VDC
Operating Humidity	$\leq 85\%$ RH

Electrode Constants and Ranges

K Value	EC Range	TDS Range
0.1 cm ⁻¹	0.2–200 µS/cm	0.1–100 ppm
1.0 cm ⁻¹	2–10,000 µS/cm	1–5,000 ppm
10.0 cm ⁻¹	20–100,000 µS/cm	10–50,000 ppm

3. Wiring

Wire Color	Function	Description
Red	V+	Power Supply (12-24V DC)
Black	GND	Power Ground
Yellow	RS485-A	Data+
Green	RS485-B	Data-

4. Communication Settings

Parameter	Value
Protocol	Modbus-RTU
Baud Rate	9600 bps
Data Bits	8
Parity	None
Stop Bits	1
Default Address	0x01

5. Register Map

Measurement Registers (Function 0x03)

Address	Description	Scaling
0x0000	Electrical Conductivity	Direct ($\mu\text{S}/\text{cm}$)
0x0001	Temperature	Value $\div 10$ ($^{\circ}\text{C}$)

Configuration Registers (Function 0x06)

Address	Description	Range/Notes
0x0004	Slope	Value $\times 1000$ (default: 1000)
0x000B	Intercept/Offset	Direct ($\mu\text{S}/\text{cm}$, default: 0)
0x0050	Slave Address	0x01–0xFE

6. Reading Data

Read 2 registers from 0x0000:

Request: 01 03 00 00 00 02 C4 0B
Response: 01 03 04 02 AE 01 64 5C 91

Decoding:

Register	Hex	Decimal	Scaling	Result
EC	0x02AE	686	Direct	686 $\mu\text{S}/\text{cm}$
Temperature	0x0164	356	$\div 10$	35.6$^{\circ}\text{C}$

7. Address Configuration

Change Address (0x01 → 0x02)

01 06 00 50 00 02 08 0B

Broadcast Discovery

Use address 0xFE with only one device connected.

8. Calibration

Set Slope (1.200)

01 06 00 04 04 B0 [CRC]

Set Intercept (10 μ S/cm)

01 06 00 0B 00 0A [CRC]

9. Installation Notes

Sensor Placement

- Stable water flow
- Fully submerged
- Avoid dead space

Maintenance

- Keep electrode clean
- No disassembly
- Avoid oils/grease

AGRINOVO Modular IoT sensors and controllers for agriculture and aquaculture.
Request a quote: agrinovo.io/contact · Full range: agrinovo.io/products
contact@agrinovo.io · WhatsApp **+972 54 688 8148**